

BIDDING DOCUMENTS

TERMS & CONDITION

The Following documents should be attached with the proposal.

- Proof of the registration with income tax/ sales Tax departments being tax payer name in ATL list.
- Income tax returns, Sales Tax returns.
- Financial status of the firms including bank statement.
- Past experience in Govt/semi Govt/Private.
- List of clients/Reference.
- List of Tools for subject repair.
- The Company will repair the equipment and handover in complete working condition to Ayub Teaching hospital.
- Firm must be provided "No defaulter Certificate" on stamp paper.
- The repair work will be within one year warranty.
- Bid Security of Rs.10000/- refundable.
- High value equipment will be repair in hospital premises.

Safad
Alam

TENDER LIST FOR REPAIR OF ELECTRO MEDICAL EQUIPMENT.

S.NO	EQUIPMENT	QTY	PROBLEMS/ERRORS
1.	Neuro Microscope Leica F 40	01	Software issue, Xenon lamp, calibration of unit
2.	CUSA (cavitron ultrasonic surgical aspirator)	01	Power Supply card issue, software
3.	Fluoroscopy Siemens (sireskope Germany)	01	Error Nos. 246-32, 606-80, 242-32
4.	Thoramate Chest X-ray Siemens	01	UP/Down Problem (BUCKY), Auto control system issue
5.	SMAM General X-Ray Italy	01	Error No. 21 Tube Error
6.	Portable X-Ray (MXCHT. 1300RM)	01	Key Board and Power supply Problem,
7.	Toray Fx Dialysis Machine	03	System Error, Water 1 & 2, Switch off, Line Filter, Rinsing Valve, Rinse Port Red, Rinse Port Blue, Silicon Tube
8.	ETT Machine Burdick T 500 Siemens	01	Control system
9.	Dental Units Planmica	04	Control PCB, Suction, Unit Control Problem, Water System Problem
10.	Ultra sound Machine (New Tech)	01	Noise Error
11.	ENT Operating Microscope (carl Zeiss)	02	Zooming out Problem, Malfunctioning
12.	Baby warmer (Ohmeda)	02	Power card & control system
13.	Anaesthesia Machine Candvish 680M	05	Calibration of Machine, Ventilator and vaporizers

Handwritten signature

14.	Cryo-centrifuge (Blood separator)	01	Weight balancing
15.	Getinge steam sterilizer	01	Calibration
16.	ICU Ventilator (Siemens 900)	04	Oxygen saturation issue, cells mixture and Complete repair up to Operational status.
17.	Pneumatic Drill Con Med (Orthopedic)	03	Bearing Jammed, Not operational.
18.	Cardiac Monitor	04	Screen issue, calibration needed.

Mad