

AYUB MEDICAL TEACHING INSTITUTION ABBOTTABAD

OFFICE OF A. DIRECTOR
ELECTROMEDICAL & EQUIPMENTS

sajjad.khan00@yahoo.com
Phone 0992-381846 Ex. 3303

No.ATH/EME 164-67 /18


Dated: 17-12-2018

✓
Deputy Director
Procurement Cell
AMTI, Abbottabad.

SUBJECT: BIDDING DOCUMENTS FOR REPAIR OF DENTAL EQUIPMENT

Reference to your letter No. D.D.ATH (P: CELL)/18/ 4546 Dated 12-12-2018 regarding the provision of Bidding documents for repair of Dental Equipment.

The Term & Conditions and Specification for Repair of Subject Equipment is herewith attached.


Asst. Director EM

Copy:

- Hospital Director, AMTI, Abbottabad
- Incharge Procurement Cell, AMTI, Abbottabad
- HOD Dentistry , AMTI, Abbottabad

TO, **HOSPITAL DIRECTOR**
MEDICAL INSTITUTION ABBOTTABAD.

Dated: 12-11-2018

Details of Medical equipments in different sections of dental department
Medical teaching institution Abbottabad.

01. Dental operative Department				Total number of units 12
Sr. No.	Name of Equipment	Model / Make	Fault / Status	Remarks
1	Dental unit with dental chair.	Mate	Totally out of order.	Needs complete repair along with replacement of defective parts
2	Dental unit with dental chair.	Mate	Water system not working. Needs complete general servicing and replacement of defective parts.	
3	Dental unit with dental chair.	Mate	Totally out of order.	Needs complete repair along with replacement of defective parts
4	Dental unit with dental chair.	Mate	Unit is working without suction and water system.	
5	Dental unit with dental chair.	Gloria	Totally out of order.	Needs complete repair along with replacement of defective parts
6	Dental unit with dental chair.	Mate	Unit is working, suction and water supply not working.	Needs general servicing and repair of water supply and suction.
7	Dental unit with dental chair.	Mate	Unit is totally out of order	Needs complete repair along with replacement of defective parts
8	Dental unit with dental chair.	Mate	Unit is working without light and water supply system	Needs complete repair along with replacement of defective parts
9	Dental unit with dental chair.	Mate	Unit is working without water supply and head light	Needs complete repair and General servicing.
10	Dental unit with dental chair.	Mate	Totally out of order	Needs complete repair and General servicing.
11	Dental unit with dental chair.	Mate	Totally out of order	Needs complete repair and General servicing.
12	Dental unit with dental chair.	Mate	Totally out of order	Needs complete repair and General servicing.

Total Units 12 Nos.
5 Units working partially.
7 Units are out of order.

TO, **HOSPITAL DIRECTOR**
MEDICAL INSTITUTION ABBOTTABAD.

Dated: 12-11-2018

Details of Medical equipments in different sections of dental department
Medical teaching institution Abbottabad.

01. Dental operative Department

Total number of units 12

Sr. No.	Name of Equipment	Model / Make	Fault / Status	Remarks
1	Dental unit with dental chair.	Mate	Totally out of order.	Needs complete repair along with replacement of defective parts
2	Dental unit with dental chair.	Mate	Water system not working. Needs complete general servicing and replacement of defective parts.	
3	Dental unit with dental chair.	Mate	Totally out of order.	Needs complete repair along with replacement of defective parts
4	Dental unit with dental chair.	Mate	Unit is working without suction and water system.	
5	Dental unit with dental chair.	Gloria	Totally out of order.	Needs complete repair along with replacement of defective parts
6	Dental unit with dental chair.	Mate	Unit is working, suction and water supply not working.	Needs general servicing and repair of water supply and suction.
7	Dental unit with dental chair.	Mate	Unit is totally out of order	Needs complete repair along with replacement of defective parts
8	Dental unit with dental chair.	Mate	Unit is working without light and water supply system	Needs complete repair along with replacement of defective parts
9	Dental unit with dental chair.	Mate	Unit is working without water supply and head light	Needs complete repair and General servicing.
10	Dental unit with dental chair.	Mate	Totally out of order	Needs complete repair and General servicing.
11	Dental unit with dental chair.	Mate	Totally out of order	Needs complete repair and General servicing.
12	Dental unit with dental chair.	Mate	Totally out of order	Needs complete repair and General servicing.

Total Units 12 Nos.
5 Units working partially.
7 Units are out of order.

02. Orthodontist section

Sr. No.	Name of Equipment	Model / Make	Fault / Status	Remarks
1	Dental unit with dental chair.	Mate	Totally out of order	Needs complete repair along with replacement of defective parts
2	Dental unit with dental chair.	Mate	Unit is giving electric shock and totally out of order.	Unit may not use for patient. Needs complete repair along with replacement of defective parts
3	Dental unit with dental chair.	Mate	Totally out of order.	Needs complete repair along with replacement of defective parts
4	Dental unit with dental chair.	Gloria	Totally out of order.	Needs complete repair along with replacement of defective parts
5	Dental Air Compressor	Gen Air	Compressor is faulty, leakage in head section.	Air Compressor needs complete repair and overhauling.

All Dental Units are not functioning.

03. Dental Surgery Male

Sr. No.	Name of Equipment	Model / Make	Fault / Status	Remarks
1	Dental unit with dental chair.	Bell Mount	Totally out of order	Unit has been complete 100 percent efficient life needs to be replace.
2	Dental unit with dental chair.	Mate	Totally out of order	Unit needs complete repair and maintenance along with defective parts.
3	Dental unit with dental chair.	Mate	Unit is working with out water supply and suction.	Unit needs complete repair and maintenance

Total Units 3 Nos

2 Units non functioning 1 unit is partially working

04. Dental Surgery Female

Sr. No.	Name of Equipment	Model / Make	Fault / Status	Remarks
1	Dental unit with dental chair.	Mate	Totally out of order	Unit has been complete 100 percent efficient life needs to be replace.
2	Dental unit with dental chair.	Mate	Totally out of order	Unit needs complete repair and maintenance along with defective parts.
3	Dental unit with dental chair.	Mate	Totally out of order	Unit needs complete repair and maintenance

**Total Units 3
All out of order**

Note: Surgery Department needs independent air compressor to drive the units for smooth working.

05. PerioDental Department

Sr. No.	Name of Equipment	Model / Make	Fault / Status	Remarks
1	Dental unit with dental chair.	Mate	Unit is working needs complete general servicing	Needs servicing and maintenance
2	Dental unit with dental chair.	Mate	Unit is working needs complete general servicing	Needs servicing and maintenance
3	Dental unit with dental chair.	Mate	Unit is not working properly dental chairs movement fault dental light bulb faulty.	Units needs complete repair and maintenance along with defective parts
4	Dental unit with dental chair.	Mate	Totally out of order	Units needs complete repair and maintenance along with defective parts

**Total Units 4 Nos
3 Units Partially working, 1 Unit out of order.**

06. Peads Dental Department

Sr. No.	Name of Equipment	Model / Make	Fault / Status	Remarks
1	Dental unit with dental chair.	Mate	Complete out of order	Needs servicing and maintenance along with defective parts
2	Dental unit with dental chair.	Mate	Dental Chair defective, Dental Chair Head light not proper working. Suction and water system defective.	Units needs complete repair and maintenance along with defective parts
3	Dental unit with dental chair.	Mate	Dental unit is working, Dental chair head needs replacement of halogen lamp.	Units needs complete repair and maintenance along with replacement of defective parts.

Total Units 3 Nos
1 Unit is working, 2 Units out of order.

07. Prospodentist Dental Department

Sr. No.	Name of Equipment	Model / Make	Fault / Status	Remarks
1	Dental unit with dental chair.	Bell Mount Japan	Totally out of order	Unit has been complete 100 percent efficient life needs to be replace.
2	Dental unit with dental chair.	Bell Mount Japan	Totally out of order	Unit has been complete 100 percent efficient life needs to be replace.
3	Dental unit with dental chair.	Bell Mount Japan	Totally out of order	Unit has been complete 100 percent efficient life needs to be replace.
4	Dental unit with dental chair.	Mate	Working but not properly	Needs complete repair and maintenance along with replacement of defective parts.
5	Dental unit with dental chair.	Mate	Working but not properly	Needs complete repair and maintenance along with replacement of defective parts
6	Dental unit with dental chair.	Bell Mount Japan	Working but not properly	Needs complete repair and maintenance along with replacement of defective parts
7	Dental unit with dental chair.	Bell Mount Japan	Working but not properly	Needs complete repair and maintenance along with replacement of defective parts

Total Units 7 Nos
4 Units are working partially, 3 Units are non functioning.

08. Dental X-Ray Department

Sr. No.	Name of Equipment	Model / Make	Fault / Status	Remarks
1	Dental X-ray	Premadent 665.kv.8ma	Working	Only single unit working in whole dental department
2	Dental X-ray	Bell Mount (Searcher) Model 066	Unit is completely out of order	X-ray needs complete repair along with defective parts
3	Dental X-ray	Bell Mount (Source-1) Model Dx 066	Unit is completely out of order	X-ray needs complete repair along with defective parts
4	Dental X-ray	Bell Mount Japan	X-ray Tube head has gone for repair.	Write letter to respective firm complete the repair work and reinstall the X-ray tube head
5	Dental X-ray	Bell Mount Japan	Totally out of order.	Replace the unit
6	Dental X-ray	Bell Mount Japan	Totally out of order. X-ray control is missing	Replace the unit
7	Dental X-ray	Bell Mount Japan	Totally out of order.	Replace the unit
8	Dental X-ray	Bell Mount Japan	Totally out of order.	Needs complete repair and maintenance along with replacement of defective parts

**Total No. of X-ray 8
1 is working 7 Units are non functioning.**

Note: X-ray Department needs at least 3 new X-ray units along with accessories likely lead apprans and purchase six new dental units as per work load.

09. Dental X-Ray Department

Sr. No.	Name of Equipment	Model / Make	Fault / Status	Remarks
1	OPG Machine	PantosDG Pluse	Working	Only single OPG unit working in whole dental department
2	OPG Machine	EC Proline (PLANMECA) PM 2002	Unit is totally non functional	OPG Machine has been completed 100% efficient life. Not Repairable. Replace with new one.

**Total OPG Machine 2 Nos
1 is working, 1 is out of order.**

Bidding Documents

Terms & Condition

The Following documents should be attached with the proposal.

- Proof of the registration with income tax/ sales Tax departments being tax payer name in ATL list.
- Income tax returns, Sales Tax returns.
- Financial status of the firms including bank statement for 3 years.
- Past experience in Govt/semi Govt/Private (Last 3 years minimum).
- List of clients/Reference.
- List of Tools for subject repair.
- The Company will repair the equipment and handover in complete working condition to Ayub Teaching hospital.
- Firm must be provided "No defaulter Certificate" on stamp paper.
- The repair work will be within one year warranty.
- Bid Security of 2% refundable.
- The Tender shall be single stage Single Envelope basis
- 10% performance guarantee deducted from bills up to warranty period.

Sr. No.	Name of Equipment	Make / Model	Fault / Status	Remarks
1	Auto Clave	Local Made	Not functioning	Needs urgent repair
2	Automatic sterilizers	European	Not functioning	Needs urgent repair