

TERMS AND CONDITIONS.

Sealed and postal tenders are invited for the following to Ayub Teaching Hospital Abbottabad for the Year 2021-22 from the eligible firms as per bidding documents.

1. Active Status on ATL
2. NTN Registration.
3. GST Registration.
4. Last year Income Tax.
5. Last year sale tax return.
6. Last three years Bank Statement.
7. Registration With KPRA (Khyber Pakhtunkhwa Revenue Authority).
8. Affidavit on stamp paper duly attested regarding not blacklisting by any Govt /Semi Govt: departments
9. Bid security in shape of Call Deposit @ 2 % against quoted calculated amount may attach with the bid.
10. Three years Market experience (certificate must be attached)
11. Past performance of the firm regarding the repair of dental equipment. (certificate must be attached)
12. List of tools and technical personal for relevant work.
13. Rate Quoted must be in printed form: hand written quoted rate will be rejected.
14. Tender will be received and open as per advertisement notice.
15. Tender will be considered according to KPPRA rules.
16. The KPPRA standard bidding documents/ General Condition of the contract are part of the bidding Documents
17. Tender will be single stage single envelop basis.
18. The tender envelops must be sealed and clear Marked as "Name of Tender" which shall be addressed to Hospital Director Ayub Teaching Hospital.
19. The bidding documents fee Rs. 1500/- is non-refundable.
20. Incomplete tender will not be accepted.
21. Conditional tender will not be accepted.
22. The price must be quoted as per Tender list.
23. The only single price must be quoted against each item.
24. The repair is found below the specification the consignment will be rejected by the inspection committee.
25. The competent authority can rejected the tender as per KPPRA rules.
26. The tender will be rejected straight away by the competent authority without assigning any reason if the firms declared black listed by any Govt./Semi govt. department/ agencies etc.
27. All prices mentioned in the financial proposal shall be in Pak Rupees (PKR). Any item and product found below the contract specification, the same shall be replace by the supplier at his own cost during the term of the warranty/guarantee period.



28. The quoted rates should be inclusive of all Govt taxes applicable time to time & any additional charges imposed by the Govt. Moreover the Hospital authority shall not be paid any charges on account of transportation /installation/commissioning etc.
29. Proposals shall be submitted in English language
30. Each page of the tender shall be signed by the bidder. If don't signed the bid will be rejected.
31. The Hospital authority may increase or decrease quantities of items as per KPPRA rules.
32. The firm will repair the dental unit, Dental X rays and compressor within the premises of the ATH.
33. The firm will hand over the dental unit, Dental X rays and compressor in complete working condition.
34. The firm will inspect the dental unit and Dental X Rays machine before submitting tender
35. The firm will quoted the repair price of each equipment separately.
36. The equipment will be one year warranty from the date of its repair/operational.
37. The firm will hand over the faulty parts to Electro Medical Equipment store ATH.
38. The firm will repair the equipment in presence of Equipment Technician ATH and will train him for the minor repair in equipment.
39. The firm will complete the repair work within 15 days after awarding of contract agreement.
40. The Partial repair cannot be accepted in any case.

13-9-21
HOSPITAL DIRECTOR
Ayub Teaching Hospital

THE LIST OF REQUIRED ITEMS WITH
QUANTITY IS ATTACHED HEREWITH AS
PART OF TENDER DOCUMENTS

SUMMARY OF DENTAL UNITS

Sr. No	Name of Equipment	Serial No.	Department	Manu:	Fault diagnosed	Remarks
1.	Dental Unit	Sr. 5108305	Operative Department	Mate	<ul style="list-style-type: none"> a) Power Supply control Board Faulty. b) 02 Transformers c) Up/down movement not working d) Forward and backward Movement not working 	Only Light in working.
2.	Dental Unit	Sr. 5108304	Operative Department	Mate	<ul style="list-style-type: none"> a) Power Supply control Board Faulty. b) Up/down movement not working c) 02 Transformers d) Hydraulic Oil e) Solenoid Valve f) Switches not working g) Light not working 	Completely Out of order
3.	Dental Unit	Sr. 5108306	Operative Department	Mate	<ul style="list-style-type: none"> a) Power Supply control Board Faulty. b) 02 Transformers c) Up/down movement not working d) Forward and backward Movement not working 	Only Light in working.
4.	Dental Unit	Sr. 5108302	Operative Department	Mate	<ul style="list-style-type: none"> a) Power Supply control Board Faulty. b) 02 Transformers c) Up/down movement not working d) Forward and backward Movement not working 	Only Light in working.

5.	Dental Unit	Sr. 5108310	Operative Department	Mate	<ul style="list-style-type: none"> a) Power Supply control Board Faulty. b) 02 Transformers c) Up/down movement not working d) Forward and backward Movement not working 	Only Light in working.
6.	Dental Unit	Sr. 5108301	Operative Department	Mate	<ul style="list-style-type: none"> a) Power Supply control Board Faulty. b) 02 Transformers c) Up/down movement not working d) Forward and backward Movement not working 	Only Light in working.
7.	Dental Unit	Sr. 5108303	Operative Department	Mate	Needs service	Functional
8.	Dental Unit	Sr. 5408310	Operative Department	Mate	<ul style="list-style-type: none"> a) Power Supply control Board Faulty. b) Up/down movement not working c) Forward and backward Movement not working d) 02 Transformers e) Hydraulic Oil f) Solenoid Valve g) Switches not working h) Light not working 	Completely Out of order
9.	Dental Unit	Sr. 5408309	Operative Department	Mate	<ul style="list-style-type: none"> a) Power Supply control Board Faulty. b) Up/down movement not working c) Forward and backward Movement not working d) 02 Transformers e) Hydraulic Oil f) Solenoid Valve 	Completely Out of order

					g) Switches not working Light not working	
10.	Dental Unit	Sr. 5110304	Ortho Department	Mate	a) Power Supply control Board Faulty. b) Up/down movement not working c) Forward and backward Movement not working d) 02 Transformers e) Hydraulic Oil f) Solenoid Valve g) Switches not working Light not working	Completely Out of order
11.	Dental Unit	Sr. 5110306	Ortho Department	Mate	a) Power Supply control Board Faulty. b) Up/down movement not working c) Forward and backward Movement not working d) 02 Transformers e) Hydraulic Oil f) Solenoid Valve g) Switches not working h) Light not working	Completely Out of order
12.	Dental Unit	Sr. 5110309	Ortho Department	Mate	a) Power Supply control Board Faulty. b) Up/down movement not working c) Forward and backward Movement not working d) 02 Transformers e) Hydraulic Oil f) Solenoid Valve g) Switches not working	Completely Out of order

					h) Light not working	
13.	Dental Unit	Sr.5110308	Female surgery	Mate	<ul style="list-style-type: none"> a) Power Supply control Board Faulty. b) Up/down movement not working c) Forward and backward Movement not working d) 02 Transformers e) Hydraulic Oil f) Solenoid Valve g) Switches not working h) Light Cover not existing. i) Light not working 	Completely Out of order
14.	Dental Unit	Sr. 5110307	Female surgery	Mate	<ul style="list-style-type: none"> a) Power Supply control Board Faulty. b) Up/down movement not working. c) Forward and backward Movement not working d) 02 Transformers e) Hydraulic Oil f) Solenoid Valve g) Switches not working h) Light broken 	Completely Out of order
15.	Dental Unit	Sr. 5110305		Mate	<ul style="list-style-type: none"> a) Power Supply control Board Faulty. b) 02 Transformers c) Up/down movement not working d) Forward and backward Movement not working 	Only Light in working.

16.	Dental Unit	Sr. 5406204	Male Surgery	Mate	<ul style="list-style-type: none"> a) Power Supply control Board Faulty. b) Up/down movement not working c) Forward and backward Movement not working d) 02 Transformers e) Hydraulic Oil f) Solenoid Valve g) Switches not working h) Light not working 	Completely Out of order
17.	Dental Unit	Sr. 5408312	Male Surgery	Mate	<ul style="list-style-type: none"> a) Power Supply control Board Faulty. b) Up/down movement not working c) Forward and backward Movement not working d) 02 Transformers e) Hydraulic Oil f) Solenoid Valve g) Switches not working h) Light not working 	Completely Out of order
18.	Dental Unit	Sr. 5408302	Periodontonic	Mate	<ul style="list-style-type: none"> a) Totally out of order b) Input power not working c) Unit short circuited d) Compressor cover (casing) out of order 	Completely Out of order
19.	Dental Unit	Sr. 5408304	Periodontonic	Mate	<ul style="list-style-type: none"> a) Totally out of order b) Input power not working c) Unit short circuited d) Compressor cover (casing) out of order 	Completely Out of order
20.	Dental Unit	Sr. 5408305	Periodontonic	Mate	<ul style="list-style-type: none"> a) Totally out of order b) Input power not working 	Completely Out of order

					c) Unit short circuited d) Compressor cover (casing) out of order	
21.	Dental Unit	Sr. 5408303	Periodontonic	Mate	a) Totally out of order b) Input power not working c) Unit short circuited d) Compressor cover (casing) out of order	Completely Out of order
22.	Dental Unit	Sr. 5110303	Room No13	Mate	a) Power Supply control Board Faulty. b) 02 Transformers c) Up/down, Forward/backward movement not working d) Movement not working	Only Light in working.
23.	Dental Unit	Sr. 5110302	Room No13	Mate	a) Power Supply control Board Faulty. b) 02 Transformers c) Up/down movement not working d) Forward and backward Movement not working	Only Light in working.
24.	Dental Unit	Sr. 5110301	Room No13	Mate	a) Power Supply control Board Faulty. b) Up/down, forward/back movement not working c) 02 Transformers d) Switches not working e) Light not working	Completely Out of order
25.	Dental Unit	Sr. 5408301	Clinical Lab	Mate	a) Power Supply control Board Faulty. b) Up/down, forward/back movement not working c) 02 Transformers d) Switches not working e) Light not working	Completely Out of order

26.	Dental Unit	Sr. 5408307	Clinical Lab	Mate	<ul style="list-style-type: none"> a) Power Supply control Board Faulty. b) Up/down, forward/back movement not working c) 02 Transformers d) Switches not working e) Light not working 	Completely Out of order
27.	Dental Unit	Sr. 5110311	Operative Department	Mate	<ul style="list-style-type: none"> a) Power Supply control Board Faulty. b) Up/down, forward/back movement not working c) 02 Transformers d) Switches not working e) Light not working 	Completely Out of order
28.	Dental Unit	Sr. 5408311	Operative Department	Mate	<ul style="list-style-type: none"> a) Power Supply control Board Faulty. b) Up/down, forward/back movement not working c) 02 Transformers d) Switches not working e) Light not working 	Completely Out of order
29.	Dental Unit	Sr. 511172LX	Clinical Lab	Belmont	<ul style="list-style-type: none"> a) Power Supply control Board Faulty. b) Up/down, forward/back movement not working c) 02 Transformers d) Switches not working e) Light not working 	Completely Out of order
30.	Dental Unit	FBk 0649081	Clinical Lab	Belmont	<ul style="list-style-type: none"> a) Power Supply control Board Faulty. b) Up/down movement not working c) 02 Transformers 	Only Light in working.

31.	Dental Unit	FBJ 0446558	Clinical Lab Room 2	Belmont	<ul style="list-style-type: none"> a) Power Supply control Board Faulty. b) Up/down, forward/back movement not working c) 02 Transformers d) Switches not working e) Light not working 	Completely Out of order
32.	Dental Unit	Sr. J0446560	Male Surgery	Belmont	<ul style="list-style-type: none"> a) Power Supply control Board Faulty. b) Up/down, forward/back movement not working c) 02 Transformers d) Switches not working e) Light not working 	Completely Out of order
33.	Dental Unit	Sr. J0446558	Clinical Room 2	Belmont	<ul style="list-style-type: none"> a) Up/down, forward back movement b) Transformers c) Switches not working d) Light not working 	Completely Out of order
34.	Dental Unit	Sr. J0847146	Clinical Room 2	Belmont		Functional
35.	Dental Unit	Sr. 316699	Surgery 1	Planmeca	<ul style="list-style-type: none"> a) Power Supply control Board Faulty. b) Up/down, forward/back movement c) 02 Transformers d) Switches not working e) No suction 	Light working but not focusing
36.	Dental Unit	Sr. 316769	Surgery 1	Planmeca	<ul style="list-style-type: none"> a) Input Power Supply not working. b) Power Supply control Board Faulty. c) Up/down, forward/back movement d) 02 Transformers 	Completely Out of order

					<ul style="list-style-type: none"> e) Switches not working f) No suction 	
37.	Dental Unit	Sr. 316695	Surgery 1	Planmeca	<ul style="list-style-type: none"> a) Power Supply control Board Faulty. b) Up/down, forward/back movement c) 02 Transformers d) Light not working e) No suction 	Completely Out of order
38.	Dental Unit	5408313	Surgery HOD	Planmeca	<ul style="list-style-type: none"> a) Power Supply control Board Faulty. b) Up/down, forward/back movement not working c) 02 Transformers 	Only Light in working.
39.	Dental Unit	540831	Operative Department	Gloria	<ul style="list-style-type: none"> a) Input Power Supply not working. b) Power Supply control Board Faulty. c) Up/down, forward/back movement d) 02 Transformers e) Switches not working f) No suction 	Completely Out of order

40.	Dental Unit	5408314	HOD Ortho Dr. Sher shah	Gloria	a) Power Supply control Board Faulty. b) Up/down, forward/Back movement not working c) 02 Transformers	Only Light in working.
41.	Dental Unit	5406203	HOD Orthodontics Dr. Fiaz	Gloria	a) Power Supply control Board Faulty. b) Up/down, forward/Back movement not working c) 02 Transformers	Only Light in working.

SUMMARY OF DENTAL X-RAYS AND OPG MACHINE

SR. NO	NAME OF EQUIPMENT	SERIAL NO.	MAKE	FAULT DIAGNOSED	REMARKS
1	Dental X-Ray	Sr. 72008	Belmont	X-Ray Tube and Control Board faulty	Replacement of X-ray Tube and repair of control system
2	Dental X-Ray	Sr.J5024	Belmont		
3	Dental X-Ray	Sr. J5006	Belmont		
4	Dental X-Ray	Sr. J05067	Belmont		
5	Dental X-Ray	Sr. J05008	Belmont		
6	Dental X-Ray	Sr. J09059	Belmont		
7	Dental X-Ray	Sr. CIF402619	EC Proline		
8	Dental X-Ray	Sr. N12874	PM 2002		
	OPG machine		Italy	Main UPS and Software of the system has been damaged	Installation of software and repair of UPS system.