

TECHNICAL SPECIFICATION OF EQUIPMENT FOR AYUB TEACHING HOSPITAL.

1. 3 CHANNEL ECG MACHINES.

- a. Integrated rechargeable battery for at least 120min or more of normal use.
- b. ECG interpretation software consistent with current/latest STEMI guidelines.
- c. 5 inch or more color LCD screen for viewing traces and input of patients details
- d. QWERTY Pad Main and battery operation
- e. Tabs electrode option
- f. Advance filter for baseline electrical and muscle tremor interference
- g. Copy print out option-automatic and manual mode option.
- h. Print review capability option if possible.
- i. Defibrillation protection thermal recorder for printout and pc print out option and ECG
- j. Storage of 50 ECG OR MORE.
- k. Z folds ECG paper/ ECG Roll